

July 10, 2024

PHILIPS RESPIRONICS

1001 Murry Ridge Lane
Murrysville, PA 15668

RE: Airline Travel with DreamStation 2 devices

To whom it may concern,

The following Philips DreamStation 2 devices are in compliance with commercial airline EMI/RFI requirements:

- DreamStation 2 CPAP
- DreamStation 2 Auto CPAP
- DreamStation 2 CPAP Advanced
- DreamStation 2 Auto CPAP Advanced

Philips has designed and tested the identified devices for compliance with section 21, Category M, RTCA DO-160 EMI/RFI requirements as specified in the Code of Federal Regulations 14 CFR 382 “*Nondiscrimination on the Basis of Disability in Air Travel; Final Rule*”.

In accordance with these requirements, the aforementioned components may be used onboard an aircraft without further testing by the carrier. The device is suitable for use on airlines when the device is operating from an AC or DC power source. **Important Note:** Do not travel with water in the water tank.

The following symbol is located in the User Manual and on the bottom of the device to indicate that it is medical equipment suitable for airline use:



If you have any other questions regarding our products, please feel free to call the Philips Respironics Customer Services department 1-724-387-4000. You can also use the following address:

Philips Respironics
1001 Murry Ridge Lane
Murrysville, PA 15668

Karlee Criado

Karlee Criado
Principal Regulatory Affairs Engineer
Philips Sleep & Respiratory Care